Schweizerische Eidgenossenschaft
Confédération suisse
Confederazione Svizzera
Confederaziun svizra

Swiss Confederation

Federal Department of Economic Affairs,
Education and Research EAER

State Secretariat for Economic Affairs SECO
Swiss Accreditation Service SAS

SCS Directory

Accreditation number: SCS 0108

International standard:	ISO/IEC 17025:2017
Swiss standard:	SN EN ISO/IEC 17025:2018

Schmid & Partner Engineering AG
Zeughausstrasse 43
8004 Zürich

Head:	Dr Sven Kühn
Responsible for MS:	Prof. Dr Niels Kuster
Telephone:	+41 44 245 97 00
E-Mail:	info@speag.swiss
Internet:	www.speag.swiss
Initial accreditation:	17.09.2004
Current accreditation:	17.09.2024 to 16.09.2029
Scope of accreditation see:	www.sas.admin.ch (Accredited bodies)

Scope of accreditation as of 18.05.2026

Calibration Laboratory for Electric Field, Magnetic Field and SAR Sensors and Dielectric Measurement Instrumentation

Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Electric field				
Calibration of E-field probes	0.8 V/m … 800 V/m	4 MHz … 6 GHz	5,1 %	e.g. ER3DV6x, EF3DVx, EU2DVx, EE3DVx, EL3DVx
Calibration of E-field probes	10 V/m … 2000 V/m	750 MHz … 6 GHz 6 GHz … 110 GHz	5,1 % 0,98 dB	e.g. EUmmWVx
Calibration of E-field probes	2 V/m … 420 V/m	3 kHz … 10 MHz	1,06 dB	e.g.,MAGPy-8H3D+E3D
Magnetic field				
Calibration of H-field probes	2 mA/m … 2 A/m	4 MHz … 3 GHz	5,1 %	e.g. H2DVx, H3DVx, HL3DVx
Calibration of H-field probes	10 A/m … 2200 A/m	3 kHz … 10 MHz	0,70 dB	e.g., MAGPy-8H3D+E3D

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Calibration of sensitivity for magnetic field probes in the audio range	0,001 … 0,1 V/(A/m)	1 kHz 0,1 … 1 A/m	2,2 %	e.g. AM1DVx
Calibration of magnetic field simulator	-30 … +40 dB A/m	1 kHz	4,1 %	e.g. TMFS (Telephone Magnetic Field Simulator)
Calibration of magnetic field simulators for validation and verification	2.5 A/m … 800 A/m	3 kHz … 10 MHz	1.13 dB	e.g., V-Coil500/3, V-Coil350/85, V-Coil50/400, V-Coil50/6780
Specific absorption rate (SAR)				e.g. ET3Dvx, ES3Dvx, EX3Dvx, ET1Dvx, EU2Dvx
	E* field (typical[2])			
Calibration of dosimetric E-field probes	0,5 V/m … 500 V/m	4 MHz … 450 MHz	6,7 % (13,3 % for SAR)	Temperature transfer calibration *) As example, the indicated range corresponds to 0,2 mW/kg … 200 W/kg for head tissue simulating liquid and f = 450 MHz
	E* field (typical[2])			e.g. ET3Dvx, ES3Dvx, EX3Dvx, ET1Dvx, EU2Dvx
Calibration of dosimetric E-field probes	0,45 V/m … 450 V/m	750 MHz … 3 GHz	5,5 % (11 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg … 200 W/kg for head tissue simulating liquid and f = 1800 MHz
	0,4 V/m … 450 V/m	3 GHz … 6 GHz	6,5 % (13,1 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg … 200 W/kg for head tissue simulating liquid and f = 5200 MHz

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
	0,4 V/m … 400 V/m	6 GHz … 10 GHz	9,3 % (18,6 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg … 200 W/kg for head tissue simulating liquid and f = 7 GHz
Calibration of temperature SAR probes	0 °C … + 60 °C	Tissue simulating Liquids	0,15 K (5 % temperature gradient for SAR)	As example, the temperature gradient measured with T1Vx and T1V3LAB probes can be determined to 5 %, which is also a contribution to SAR accuracy. (Noise is dominating the lower SAR threshold to typically 0,2 W/kg)
Calibration of test system validation sources	SAR 1 g and 10 g per 1 W input power	4 MHz … 290 MHz	18,4 % for SAR 1 g 18,0 % for SAR 10 g	e.g. CLA-6, CLA-13, CLA-30, CLA-64, CLA-128, CLA-150, CLA-220 according to IEC/IEEE 62209-1528 for 1 g and 10 g SAR
		300 MHz … 700 MHz	18,1 % for SAR 1 g 17,6 % for SAR 10 g	e.g. D835V2 … D3000V2 according to IEC/IEEE 62209-1528, for 1 g and 10 g SAR
		700 MHz … 3 GHz	17,0 % for SAR 1 g 16,5 % for SAR 10 g	
	SAR 1 g and 10 g per 1 W input power	3 GHz … 6 GHz	19,9 % for SAR 1 g 19,5 % for SAR 10 g	e.g. D3500V2 … D5GHzV2 according to IEC/IEEE 62209-1528, for 1 g and 10 g SAR
	SAR 1 g and 10 g per 1 W input power	6 GHz … 10 GHz	24,7 % for SAR 1 g 24,4 % for SAR 10 g	e.g. D6.5GHzV2 … D9GHzV2 according to IEC/IEEE 62209-1528, for 1 g and 10 g SAR
Near-Field Calibration of dipoles in air	E field per 0,1 W input power 30 V/m … 300 V/m	700 MHz … 6 GHz	9,5 % for E field	e.g. CD835V3 … CD5500V3 according to ANSI C 63.19, for E field and H field
	H field per 0,1 W input power 0,07 A/m … 0,7 A/m	385 MHz – 3 GHz	8,3 % for H field	e.g. AdxxxV5 according to ANSI 14117 for H-field

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Calibration of Golden Validation Device	SAR 1 g and 10 g per GVD input power	900 MHz	15,0 % for SAR 1 g 15,1 % for SAR 10 g	e.g. GVD – SR 004001 AA
DC Voltage				
Calibration of readout units for field and SAR probes	2 mV		1,5 %	e.g. DAE3Vx, DAE4Vx, DAEasyVx
	200 mV		1 %	
	1 mV … 500 mV		1,16 %	e.g. EASY6-DAE, DAE4ICEy
Calibration of SAR for planar array systems	SAR peak at 4 mm depth per 1 W input power	650 MHz … 3 GHz	25,8 % for iSAR Flat 29,3 % for iSAR Head	e.g. iSAR Flat, iSAR Head
	SAR peak at 3 mm depth per 1 W input power	650 MHz … 3 GHz	22,4 % for cSAR3D Flat 25,9 % for cSAR3D Left/Right Head	e.g., cSAR3D Flat, cSAR3D Left Head, cSAR3D Right Head
	SAR peak at 3 mm depth per 1 W input power	3 GHz … 6 GHz	25,1 % for cSAR3D Flat 28,3 % for cSAR3D Left/Right Head	e.g., cSAR3D Flat, cSAR3D Left Head, cSAR3D Right Head
	SAR peak at 3 mm depth per 1 W input power	6 GHz … 10 GHz	29,8 % for cSAR3D Flat	e.g. cSAR3D Flat, cSAR3D Quad
Calibration of antenna sources for transfer calibration of planar array systems	SAR peak per 1 W input power	650 MHz … 3 GHz	17,7 % for SAR peak	e.g., SA AAE 083B, SA AAE 190 A, etc.
	SAR peak per 1 W input power	3 GHz … 6 GHz	19,9 % for SAR peak	
Calibration of thermometers	0 °C … + 60 °C		0,1K	
Calibration of active electro-optical E&H-field probes optimized for close near-field evaluations in air	15 mV/m … 75 V/m	50 MHz … 6000 MHz	3,3 dB	for antenna factor (E-field)
	42 µA/m … 0.5 A/m	50 MHz … 6000 MHz	3,3 dB	for antenna factor (H-field)
Calibration of active electro-optical RF over fiber systems	Gain: -55 … 15dB	10 MHz … 10 GHz	2,0 dB	For transfer function (gain)

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Calibration of stabilized RF power sources	- 5 dBm … + 17 dBm	600 MHz … 6 GHz	0,43 dB	e.g. Powersource1
Calibration Procedure for sources in air above 6 GHz	10 V/m …2000 V/m	10 GHz… 110 GHz	1,27 dB 1,47 dB 1,54 dB	e.g. verification source e.g. validation source e.g. general source
Calibration Procedure for sources in air above 6 GHz, for H and averaged S	25 mA/m … 2,5 A/m 2.5 W/m^2…2,5 kW/m^2	10 GHz… 110 GHz	1,28 dB 1,63 dB 1,70 dB	e.g. verification source e.g. validation source e.g. general source
Absorbed Power Density (APD)				
Calibration of dosimetric E-field probes for APD measurement	20 V/m … 550 V/m	10 GHz … 45 GHz	±0.88dB (k=2)	e.g., EUAPDVx
Calibration of APD system verification sources	1.0 W/m^2 … 800 W/m^2	10 GHz … 45 GHz	±1.28dB, pAPD ±1.33dB, psAPD1cm^2 ±1.38dB, psAPD4cm^2	e.g., APD verification sources
Measurement capability of the dielectric probe	Permittivity, Conductivity or Loss Tangent from	homogeneous isotropic material		Open-ended coaxial probes, e.g. DAK-12, DAK-3.5, DAK-1.2 E
Permittivity	1…15			
		4 MHz … 20 MHz	24,3 %	Loss tangent: < 0,1
		20 MHz … 200 MHz	11,2 %	
		200 MHz … 3 GHz	2,0 %	
		3 GHz … 6 GHz	2,0 %	
		6 GHz … 20 GHz	2,1 %	
		20 GHz … 40 GHz	3,2 %	
		40 GHz … 67 GHz	3,2 %	
		40 GHz … 67 GHz	4.5%	Conductivity: > 10 S/m
	10…40			
		4 MHz … 10 MHz	6,4 %	Conductivity: < 0,1 S/m
		10 MHz … 50 MHz	3,8 %	Conductivity: < 0,1 S/m
		50 MHz … 200 MHz	1,8 %	Conductivity: 0,1 – 10 S/m
		200 MHz … 3 GHz	1,8 %	Conductivity: 0,1 – 10 S/m
		3 GHz … 6 GHz	2,3 %	Conductivity: 1 – 10 S/m
		6 GHz … 20 GHz	3,7 %	Conductivity: > 10 S/m
		20 GHz … 40 GHz	4,8 %	Conductivity: > 10 S/m
		40 GHz … 67 GHz	6,4 %	Conductivity: > 10 S/m

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
	35…100			
		4 MHz … 10 MHz	6,7 %	Conductivity: 0,1 – 1 S/m
		10 MHz … 50 MHz	2,9 %	Conductivity: 0,1 – 1 S/m
		50 MHz … 200 MHz	2,2 %	Conductivity: 1 – 10 S/m
		200 MHz … 3 GHz	1,7 %	Conductivity: 1 – 10 S/m
		3 GHz … 6 GHz	1,9 %	Conductivity: > 10 S/m
		6 GHz … 20 GHz	2,4 %	
Conductivity	1 … 10 S/m			
		4 MHz … 10 MHz	3,1 %	Permittivity: 35 – 100
		10 MHz … 50 MHz	2,9 %	Permittivity: 35 – 100
		50 MHz … 200 MHz	2,5 %	Permittivity: 35 – 100
		200 MHz … 3 GHz	3,2 %	Permittivity: 35 – 100
		3 GHz … 6 GHz	3,0 %	Permittivity: 35 – 100
		6 GHz … 20 GHz	3,0 %	Permittivity: 10 – 40
	>10 S/m	20 GHz … 40 GHz	3,8 %	Permittivity: 10 – 40
		40 GHz … 67 GHz	4,7 %	Permittivity: 1 – 15
Loss Tangent	0 … 0,1			
		4 MHz … 20 MHz	0,46	Permittivity: 1 – 15
		20 MHz … 200 MHz	0,28	
		200 MHz … 3 GHz	0,03	
		3 GHz … 6 GHz	0,03	
		6 GHz … 20 GHz	0,03	
		20 GHz … 40 GHz	0,03	
		40 GHz … 67 GHz	0,03	
Measurement capability of the dielectric probe for liquids and gels	Permittivity and Conductivity from	homogeneous isotropic material		Open-ended coaxial probes, e.g. DAK-12, DAK-3.5, DAK-1.2 E
Permittivity	1…100			
		4 MHz … 10 MHz	6,4 %	static conductivity < 0,1 S/m
		10 MHz … 20 MHz	3,8 %	
		20 MHz … 30 MHz	2,3 %	
		30 MHz … 50 MHz	1,9 %	
		50 MHz … 5 GHz	2,2 %	
		5 GHz … 20 GHz	3,7 %	
	1…100			static conductivity 0,1 – 10 S/m
		4 MHz … 10 MHz	6,8 %	
		10 MHz … 20 MHz	3,2 %	
		20 MHz … 30 MHz	3,3 %	
		30 MHz … 50 MHz	3,3 %	
		50 MHz … 5 GHz	3,1 %	
		5 GHz … 20 GHz	3,9 %	
		20 GHz … 40 GHz	5,0 %	
		40 GHz … 67 GHz	6,5 %	

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)		Remarks
Conductivity	0.01 … 0,1 S/m				Permittivity: 1 – 100
		300 MHz … 500 MHz	7,5 %		
		500 MHz … 5 GHz	4,4 %		
		5 GHz … 20 GHz	5,4 %		
Conductivity	0,1 – 10 S/m				Permittivity: 1 – 100
		4 MHz … 10 MHz	3,9 %		
		10 MHz … 20 MHz	3,9 %		
		20 MHz … 30 MHz	3,8 %		
		30 MHz … 50 MHz	3,4 %		
		50 MHz … 5 GHz	4,8 %		
		5 GHz … 20 GHz	4,8 %		
		20 GHz … 40 GHz	4,8 %		
		40 GHz … 67 GHz	4,9 %		
Capability of the dielectric probe to measure thin layers of solids and liquids in small volume	Permittivity, Conductivity or Loss Tangent from	Homogeneous isotropic material	Permittivity	Loss tangent	Open-ended coaxial probes, e.g. DAK12-TL2, DAK3.5-TL2, DAK1.2E-TL2
Permittivity Loss tangent	1< Permittivity <10 Loss tangent <0,05	4 MHz … 20 MHz	---	---	0.1<thickness<0.2 mm
		20 MHz … 30 MHz	---	---	
		30 MHz … 50 MHz	---	---	
		50 MHz … 100 MHz	---	---	
		100 MHz … 600 MHz	32,6 %	0,06	
		600 MHz … 3 GHz	29,5 %	0,03	
		3 GHz … 6 GHz	12,6 %	0,03	
		6 GHz … 20 GHz	10,0 %	0,03	
		20 GHz … 40 GHz	9,1 %	0,03	
		40 GHz … 67 GHz	4,5 %	0,03	
	1< Permittivity <10 Loss tangent <0,05	4 MHz … 20 MHz	34,6 %	0,45	0.2<thickness<1 mm
		20 MHz … 30 MHz	27,0 %	0,27	
		30 MHz … 50 MHz	25,6 %	0,17	
		50 MHz … 100 MHz	20,7 %	0,10	
		100 MHz … 600 MHz	9,1 %	0,06	
		600 MHz … 3 GHz	6,5 %	0,03	
		3 GHz … 6 GHz	3,7 %	0,03	
		6 GHz … 20 GHz	3,3 %	0,03	
		20 GHz … 40 GHz	3,9 %	0,03	
		40 GHz … 67 GHz	3,5 %	0,03	
	1< Permittivity <10 Loss tangent <0,05	4 MHz … 20 MHz	24,3 %	0,45	1<thickness<10 mm
		20 MHz … 30 MHz	11,2 %	0,27	
		30 MHz … 50 MHz	7,1%	0,17	
		50 MHz … 100 MHz	4,7 %	0,10	
		100 MHz … 600 MHz	2,7 %	0,06	
		600 MHz … 3 GHz	2,1 %	0,03	
		3 GHz … 6 GHz	2,0 %	0,03	
		6 GHz … 20 GHz	2,2 %	0,03	
		20 GHz … 40 GHz	3,9 %	0,03	
		40 GHz … 67 GHz	3,2 %	0,03	

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)		Remarks
	1< Permittivity <10 Loss tangent >0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	--- --- --- --- 18,6 % 24,6 % 14,7 % 8,3 % 10,4 % 4,7 %	--- --- --- --- 0,06 0,03 0,03 0,03 0,03 0,05	0.1<thickness<0.2 mm
	1< Permittivity <10 Loss tangent >0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	--- --- --- --- 29,8 % 19,5 % 8,0 % 4,1 % 4,3 % 3,8 %	--- --- --- --- 0,06 0,03 0,03 0,03 0,03 0,03	0.2<thickness<1 mm
	1< Permittivity <10 Loss tangent >0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	24,3 % 11,2 % 7,1% 4,7 % 2,6 % 2,0 % 1,9 % 2,0 % 3,2 % 3,1 %	0,45 0,27 0,17 0,10 0,06 0,03 0,03 0,03 0,03 0,03	1<thickness<10 mm
	Permittivity >10 Loss tangent <0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	28,4 % 18,5 % 12,6 % 8,6 % 5,7 % 5,7 % 5,7 % 4,1 % 4,6 % 6,2 %	0,45 0,27 0,17 0,10 0,06 0,03 0,03 0,03 0,03 0,03	0.1<thickness<0.2 mm
	Permittivity >10 Loss tangent <0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	24,7 % 12,1 % 8,5 % 6,6 % 3,7 % 4,0 % 3,0 % 3,5 % 3,8 % 3,9 %	0,45 0,27 0,17 0,10 0,06 0,03 0,03 0,03 0,03 0,03	0.2<thickness<1 mm

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)		Remarks
	Permittivity >10 Loss tangent <0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	24,3 % 11,2 % 7,1% 4,7 % 2,6 % 1,9 % 1,9 % 2,0 % 4,5 % 3,6 %	0,45 0,27 0,17 0,10 0,06 0,03 0,03 0,03 0,03 0,03	1<thickness<10 mm
	Permittivity >10 Loss tangent >0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	20,9 % 20,4 % 15,3 % 11,2 % 7,9 % 7,2 % 5,3 % 4,2 % 6,4 % 8,9 %	0,35 0,35 0,35 0,25 0,11 0,02 0,03 0,05 0,06 0,14	0.1<thickness<0.2 mm
	Permittivity >10 Loss tangent >0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	7,6 % 6,1 % 6,1 % 6,0 % 4,8 % 5,4 % 3,5 % 3,8 % 5,2 % 6,0 %	0,35 0,35 0,35 0,25 0,11 0,02 0,03 0,05 0,06 0,08	0.2<thickness<1 mm
	Permittivity >10 Loss tangent >0,05	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	5,2 % 2,6 % 2,6 % 2,5 % 2,5 % 2,8 % 2,8 % 3,6 % 3,3 % 4,5 %	0,35 0,35 0,35 0,25 0,11 0,02 0,03 0,05 0,06 0,08	1<thickness<10 mm

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.

SCS Directory

Accreditation number: SCS 0108

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)		Remarks
			Permittivity	Conduc-tivity	
Permittivity Conductivity	Permittivity >10 Conductivity >0,5 S/m	4 MHz … 20 MHz 20 MHz … 30 MHz 30 MHz … 50 MHz 50 MHz … 100 MHz 100 MHz … 600 MHz 600 MHz … 3 GHz 3 GHz … 6 GHz 6 GHz … 20 GHz 20 GHz … 40 GHz 40 GHz … 67 GHz	6,4 % 4,3 % 4,0 % 2,5 % 2,9 % 2,8 % 2,8 % 3,6 % 3,9 % 5,1 %	3,9 % 3,4 % 3,4 % 3,4 % 3,4 % 5,8 % 4,0 % 4,0 % 4,1 % 5,0 %	1<thickness<10 mm
Measurement capa-bility of low-loss thin materials	Permittivity <50	10 GHz … 45 GHz (3)	1% (all resonance frequen-cies)		Sample thickness range is 0.05 - 3 mm
	Loss tangent <0.01		1.0×10^{-4} (all reso-nance fre-quencies)		The measurement method is based on the split cylinder resonator

(2) Slightly depending on the frequency and probe type.

(3) Typical resonance frequencies of closed empty DAK-R: 10, 17, 26, 35 and 45 GHz

In case of contradictions in the language versions of the directories, the German version shall apply.

* / * / * / * / *

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor k = 2, which corresponds to a confidence level of about 95% for a normal distribution.