



SCS Directory

Accreditation number: SCS 0108

International standard: ISO/IEC 17025:2005
Swiss standard: SN EN ISO/IEC 17025:2005

Schmid & Partner Engineering AG
Zeughausstrasse 43
8004 Zürich
(Switzerland)

Head: Dr Katja Poković
Responsible for MS: Prof. Dr Niels Kuster
Telephone: +41 44 245 97 00
E-Mail: info@speag.swiss
Internet: <http://www.speag.swiss>
Initial accreditation: 17.09.2004
Current accreditation: 17.09.2014 to 16.09.2019
Scope of accreditation see: www.sas.admin.ch
(Accredited bodies)

Site under the accreditation:

R1308, Kranztechno, 388,
Dunchondaero
Joongwon-gu Seongnam-city,
Gyunggi-do
Korea, 462-729

Head: Jinee Park
Responsible for MS: Prof. Dr Niels Kuster
Telephone: +82 31 777 8453
E-Mail: info@speag-calkorea.com
Internet: <http://www.speag-calkorea.com>

Scope of accreditation as of 04.12.2018

Calibration laboratory for Specific Electric and Magnetic RF Fields and SAR measurements

Calibration and Measurement Capability (CMC)

Measured Quantity / Instrument or Gauge	Measurement Range	Measurement Conditions	Best Measurement Uncertainty CMC at (22 ± 3) °C (1)	Remarks
Electric field				
Calibration of E-field probes	0.8 V/m ... 800 V/m	4 MHz ... 6 GHz	5,1 %	e.g. ER3DV6x, EF3DVx, EU2DVx, EE3DVx, EL3DVx
Calibration of E-field probes	10 V/m ... 2000 V/m	750 MHz ... 6 GHz 6 GHz ... 110 GHz	5,1 % 0,98 dB	e.g. EUmmWVx



SCS Directory

Accreditation number: SCS 0108

Magnetic field				
Calibration of H-field probes	2 mA/m ... 2 A/m	4 MHz ... 3 GHz	5,1 %	e.g. H2DVx, H3DVx, HL3DVx
Calibration of sensitivity for magnetic field probes in the audio range	0,001 ... 0,1 V/(A/m)	1 kHz 0,1 ... 1 A/m	2,2 %	e.g. AM1DVx
Calibration of magnetic field simulator	-30 ... +40 dB A/m	1 kHz	4,1 %	e.g. TMFS (Telephone Magnetic Field Simulator)
Specific absorption rate (SAR)				e.g. ET3DVx, ES3DVx, EX3DVx, ET1DVx, EU2DVx
	E* field (typical ²)			
Calibration of dosimetric E-field probes	0,5 V/m ... 500 V/m	4 MHz ... 450 MHz	6,7 % (13,3 % for SAR)	Temperature transfer calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 450 MHz
	E* field (typical ²)			e.g. ET3DVx, ES3DVx, EX3DVx, ET1DVx, EU2DVx
Calibration of dosimetric E-field probes	0,45 V/m ... 450 V/m	750 MHz ... 3 GHz	5,5 % (11 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 1800 MHz
	E* field (typical ²)			e.g. EX3DVx, ET1DVx
	0,4 V/m ... 450 V/m	3 GHz ... 6 GHz	6,5 % (13,1 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 5200 MHz



SCS Directory

Accreditation number: SCS 0108

Calibration of temperature SAR probes	0 °C ... + 60 °C	Tissue simulating Liquids	0,15 K (5 % temperature gradient for SAR)	As example, the temperature gradient of T1Vx and T1V3LAB probes can be determined to 5 %, which is also contribution to SAR accuracy. (Noise is dominating the lower SAR threshold to typically 0,2 W/kg)
Calibration of test system validation sources	SAR* 1 g and 10 g per 1 W input power	4 MHz ... 290 MHz	18,4 % for SAR 1 g 18,0 % for SAR 10 g	e.g. CLA-6, CLA-13, CLA-30, CLA-64, CLA-128, CLA-150, CLA-220
		300 MHz ... 450 MHz	18,1 % for SAR 1 g 17,6 % for SAR 10 g	e.g. D835V2 ... D3000V2 according to IEEE 1528, IEC 62209-1/2, for 1 g and 10 g SAR
		750 MHz ... 3 GHz	17,0 % for SAR 1 g 16,5 % for SAR 10 g	*) SAR given (as example) for head tissue simulating liquid
Near-Field Calibration of dipoles in air	E* field per 0,1 W input power 30 V/m ... 300 V/m	700 MHz ... 6 GHz	9,5 % for E field	e.g. CD835V3 ... CD5500V3 according to ANSI C 63.19, for E field and H field
		H* field per 0,1 W input power 0,07 A/m ... 0,7 A/m	385 MHz – 3 GHz	8,3 % for H field
Calibration of Golden Validation Device	SAR* 1 g and 10 g per GVD input power	900 MHz	15,0 % for SAR 1 g 15,1 % for SAR 10 g	e.g. GVD – SR 004001 AA
DC Voltage				
Calibration of readout units for field and SAR probes	2 mV		1,5 %	e.g. DAE3Vx, DAE4Vx, DAEasyVx
	200 mV		1 %	



SCS Directory

Accreditation number: SCS 0108

Calibration of SAR for planar array systems	SAR* peak at 4 mm depth per 1 W input power	650 MHz ... 3 GHz	25,8 % for iSAR Flat 29,3 % for iSAR Head	e.g. iSAR Flat, iSAR Head
	SAR* peak at 3 mm depth per 1 W input power	650 MHz ... 3 GHz	22,4 % for cSAR3D Flat 25,9 % for cSAR3D Left/Right Head	e.g., cSAR3D Flat, cSAR3D Left Head, cSAR3D Right Head
	SAR* peak at 3 mm depth per 1 W input power	3 GHz ... 6 GHz	25,1 % for cSAR3D Flat 28,3 % for cSAR3D Left/Right Head	e.g., cSAR3D Flat, cSAR3D Left Head, cSAR3D Right Head also possible in the Subsidiary
Calibration of antenna sources for transfer calibration of planar array systems	SAR* peak per 1 W input power	650 MHz ... 3 GHz	17,7 % for SAR peak	e.g., SA AAE 083B, SA AAE 190 A, etc.
	SAR* peak per 1 W input power	3 GHz ... 6 GHz	19,9 % for SAR peak	
Calibration of thermometers	0 °C ... + 60 °C		0.1K	
Calibration of dielectric parameter probe Permittivity	Permittivity, Conductivity or Loss Tangent from 10 MHz ... 20 GHz 1...15 10...40 35...100	homogeneous isotropic material		Open-ended coaxial probes, e.g. DAK-3.5, DAK-1.2, DAK-12
		10 MHz ... 20 MHz	24,3 %	Loss tangent < 0,1
		20 MHz ... 200 MHz	11,2 %	
		200 MHz ... 3 GHz	2,0 %	
		3 GHz ... 6 GHz	2,0 %	
6 GHz ... 20 GHz	2,1 %			
	10 MHz ... 20 MHz	3,8 %	Conductivity < 0,1 S/m Conductivity < 0,1 S/m Conductivity: 1 – 10 S/m Conductivity: 1 – 10 S/m Conductivity: > 10 S/m	
	20 MHz ... 200 MHz	2,3 %		
	200 MHz ... 3 GHz	1,8 %		
	3 GHz ... 6 GHz	2,3 %		
	6 GHz ... 20 GHz	3,4 %		
	10 MHz ... 20 MHz	2,9 %	Conductivity: 0,1 – 1 S/m Conductivity: 0,1 – 1 S/m Conductivity: 1 – 10 S/m Conductivity: 1 – 10 S/m Conductivity: > 10 S/m	
	20 MHz ... 200 MHz	2,1 %		
	200 MHz ... 3 GHz	1,7 %		
	3 GHz ... 6 GHz	1,9 %		
	6 GHz ... 20 GHz	2,4 %		



SCS Directory

Accreditation number: SCS 0108

Conductivity	1 ... 10 S/m	10 MHz ... 20 MHz 20 MHz ... 200 MHz 200 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz	2,5 % 2,4 % 2,7 % 3,0 % 3,0 %	Permittivity: 35 – 100 Permittivity: 35 – 100 Permittivity:35 – 100 Permittivity:35 – 100 Permittivity:10 – 40
Loss Tangent	0 ... 0,1	10 MHz ... 20 MHz 20 MHz ... 200 MHz 200 MHz ... 3 GHz 3 GHz ... 6 GHz 6 GHz ... 20 GHz	0,46 0,28 0,03 0,03 0,03	Permittivity: 1 – 15
Calibration of dielectric parameters of liquids and gels	Permittivity and Conductivity from 10 MHz ... 50 GHz	homogeneous isotropic material		using open-ended coaxial probes, e.g. DAK-3.5, DAK-1.2, DAK-12
Permittivity	1...100	10 MHz ... 20 MHz 20 MHz ... 30 MHz 30 MHz ... 50 MHz 50 MHz ... 5 GHz 5 GHz ... 20 GHz 20 GHz ... 50 GHz	11,2 % 5,0 % 3,5 % 3,1 % 3,9 % 6,6 %	static conductivity < 0,1 S/m
	1...100	10 MHz ... 20 MHz 20 MHz ... 30 MHz 30 MHz ... 50 MHz 50 MHz ... 5 GHz 5 GHz ... 20 GHz 20 GHz ... 50 GHz	3,2 % 3,3 % 3,3 % 3,1 % 3,9 % 6,6 %	static conductivity 0,1 – 10 S/m
Conductivity	0.01 ... 0,1 S/m	300 MHz ... 500 MHz 500 MHz ... 20 GHz 20 GHz ... 50 GHz	6,5 % 5,2 % 7,9 %	Permittivity: 1 – 100
Conductivity	0,1 – 10 S/m	10 MHz ... 20 MHz 20 MHz ... 30 MHz 30 MHz ... 300 MHz 300 MHz ... 500 MHz 500 MHz ... 20 GHz 20 GHz ... 50 GHz	2,7 % 3,4 % 3,4 % 4,3 % 4,8 % 4,8 %	Permittivity: 1 – 100



SCS Directory

Accreditation number: SCS 0108

Calibration of active electro-optical E&H-field probes optimized for close near-field evaluations in air	15 mV/m ... 75 V/m	50 MHz ... 6000 MHz	3,3 dB	for antenna factor (E-field) for antenna factor (H-field)
	42 µA/m ... 0.5 A/m	50 MHz ... 6000 MHz	3,3 dB	
Calibration of active electro-optical RF over fiber systems	Gain: -55 ... 15dB	10 MHz ... 10 GHz	2,0 dB	For transfer function (gain)
Calibration of stabilized RF power sources	- 5 dBm ... + 17 dBm	600 MHz ... 6 GHz	0,43 dB	e.g. Powersource1
Calibration Procedure for sources in air above 6 GHz	10 V/m ... 2000 V/m	10 GHz... 110 GHz	1,27 dB	e.g. verification source
			1,47 dB	e.g. validation source
			1,54 dB	e.g. general source
Calibration Procedure for sources in air above 6 GHz, for averaged H and S	25 mA/m ... 2.5 A/m 2.5 W/m ² ...2.5 kW/m ²	10 GHz... 110 GHz	1,28 dB	e.g. verification source
			1,63 dB	e.g. validation source
			1,70 dB	e.g. general source

(2) Slightly depending on the frequency and probe type