



Schweizerische Eidgenossenschaft
Confédération suisse
Confederazione Svizzera
Confederaziun svizra

Federal Department of Economic Affairs DEA
State Secretariat for Economic Affairs SECO
Swiss Accreditation Service SAS

Accreditation number SCS 108
Numero d'accreditamento

Accreditation Standard ISO/IEC 17025:2005
Norma d'accreditamento ISO/IEC

SCS Directory Registro SCS

page/pagina 1 of/di 4

Calibration Laboratory for Specific Electric and Magnetic RF Fields and SAR measurements

Schmid & Partner
Engineering AG
Zeughausstrasse 43
8004 Zürich
☎ +41 44 245 97 00
Fax +41 44 245 97 79
<mailto:info@speag.com>
<http://www.speag.com>

Head of laboratory	:	Dr. Katja Poković
Deputy of head of laboratory	:	Dr. Fin Bomholt
Responsible person for MS	:	Prof. Dr. Niels Kuster
First accreditation (d,m,y)	:	17.09.2004
Last accreditation (d,m,y)	:	17.09.2009
Actual version	:	http://www.sas.ch/

Measured Quantities:

Electric field
Magnetic field

Change:

Specific Absorption Rate (SAR)
Temperature
DC Voltage

Staff	:	
Scope extension	:	01.10.05, 01.04.08, 17.02.10

Address	:	
Edition	:	SCS108/I

The given extended measurement uncertainty is the standard uncertainty of the measurement multiplied by an extension factor $k = 2$, which corresponds to a confidence level of about 95% for a normal distribution.

Measured Quantity Instrument	Range	Condition of measurement	Best Measurement Capability CMC at $(22 \pm 3)^\circ\text{C}$	Remarks
Electric field				
Calibration of E-field probes	0.8 V/m ... 800 V/m	10 MHz ... 3 GHz	4,6 %	e.g. ER3DV6x, EF3DVx, EU2DVx, EE3DVx
Magnetic field				
Calibration of H-field probes	2 mA/m ... 2 A/m	10 MHz ... 3 GHz	4,6 %	e.g. H2DVx, H3DVx
Calibration of sensitivity for magnetic field probes in the audio range	0,001 ... 0,1 V/(A/m)	1 kHz 0,1 ... 1 A/m	2,2 %	e.g. AM1DVx
Calibration of magnetic field simulator	-30 ... +40 dB A/m	1 kHz	4,1 %	e.g. TMFS (Telephone Magnetic Field Simulator)



Accreditation number SCS 108
Numero d'accreditamento

Accreditation Standard ISO/IEC 17025:2005
Norma d'accreditamento ISO/IEC

SCS Directory Registro SCS

page/pagina 2 of/di 4

Calibration Laboratory for Specific Electric and Magnetic fields and SAR measurements

Measured Quantity Instrument	Range	Condition of measurement	Best Measurement Capability CMC at $(22 \pm 3)^\circ\text{C}$	Remarks
Specific absorption rate (SAR)	E* field (typical ¹) 0,5 V/m ... 500 V/m	300 MHz ... 450 MHz	6,7 % (13,3 % for SAR)	e.g. ET3DVx, ES3DVx, EX3DVx, ET1DVx, EU2DVx Temperature transfer calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 450 MHz
Calibration of dosimetric E-field probes	E* field (typical ¹) 0,45 V/m ... 450 V/m	800 MHz ... 3 GHz	5,5 % (11 % for SAR)	Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 1800 MHz
Calibration of dosimetric E-field probes	E* field (typical ¹) 0,4 V/m ... 450 V/m	3 GHz ... 6 GHz	6,5 % (13,1 % for SAR)	e.g. EX3DVx, ET1DVx Waveguide analytical calibration *) As example, the indicated range corresponds to 0,2 mW/kg ... 200 W/kg for head tissue simulating liquid and f = 5200 MHz
Calibration of temperature SAR probes	0 °C ... + 60 °C	Tissue simulating Liquids	0,15 K (5 % temperature gradient for SAR)	As example, the temperature gradient of T1Vx and T1V3LAB probes can be determined to 5 %, which is also contribution to SAR accuracy. (Noise is dominating the lower SAR threshold to typically 0,2 W/kg)



Schweizerische Eidgenossenschaft
Confédération suisse
Confederazione Svizzera
Confederaziun svizra

Federal Department of Economic Affairs DEA
State Secretariat for Economic Affairs SECO
Swiss Accreditation Service SAS

Accreditation number SCS 108
Numero d'accreditamento

Accreditation Standard ISO/IEC 17025:2005
Norma d'accreditamento ISO/IEC

SCS Directory Registro SCS

page/pagina 3 of/di 4

Calibration Laboratory for Specific Electric and Magnetic fields and SAR measurements

Measured Quantity Instrument	Range	Condition of measurement	Best Measurement Capability CMC at $(22 \pm 3)^\circ\text{C}$	Remarks
Calibration of test system validation dipoles	SAR* 1 g and 10 g per 1 W input power SAR* 1 g and 10 g per 1 W input power	300 MHz ... 450 MHz 835 MHz ... 3 GHz 3 GHz ... 6 GHz	18,1 % for SAR 1 g 17,6 % for SAR 10 g 17,0 % for SAR 1 g 16,5 % for SAR 10 g 19,9 % for SAR 1 g 19,5 % for SAR 10 g	e.g. D835V2 ... D3000V2 according to IEEE 1528-2003, for 1 g and 10 g SAR *) SAR given (as example) for head tissue simulating liquid e.g. D3500V2 ... D5GHzV2 according to IEC 62209-2, for 1 g and 10 g SAR *) SAR given (as example) for head tissue simulating liquid
Calibration of dipoles in air	E* field per 0,1 W input power 30 V/m ... 300 V/m H* field per 0,1 W input power 0,07 A/m ... 0,7 A/m	700 MHz ... 3 GHz	12,8 % for E field 8,2 % for H field	e.g. CD835V3 ... CD2450V3 according to ANSI C 63.19 - 2007, for E field and H field
DC Voltage				
Calibration of readout units for field and SAR probes	2 mV 200 mV		1,5 % 1 %	e.g. DAE3Vx, DAE4Vx, DAEasyVx
Calibration of iSAR	SAR* peak per 1 W input power	800 MHz ... 3 GHz	24,5 % for iSAR - Flat 28,1 % for iSAR - Head	e.g. iSAR-Flat, iSAR-Head
Calibration of iSAR dipole antenna	SAR* peak per 1 W input power	800 MHz ... 3 GHz	18,1 % for SAR peak	e.g., SA AAE 083BA, SA AAE 190 AA, etc.
Calibration of dielectric probe	Permittivity, Conductivity or Loss Tangent from 20 MHz ... 6 GHz	homogeneous material		Open-ended coaxial probes



Accreditation number SCS 108
Numero d'accreditamento

SCS Directory Registro SCS

Accreditation Standard ISO/IEC 17025:2005
Norma d'accreditamento ISO/IEC

page/pagina 4 of/di 4

Measured Quantity Instrument	Range	Condition of measurement	Best Measurement Capability CMC at $(22 \pm 3)^\circ\text{C}$	Remarks
Permittivity (relative)	3...120	300 MHz ... 6 GHz	3,0 %	Liquids
		20 MHz	4,5 %	Polar liquids, e.g. Ethanol,
		30 MHz ... 6 GHz	3,2 %	Methanol, Ethanediol
		65 MHz ... 6 GHz	3,6 %	Planar low loss substrate
		100 MHz	5,0 %	Tissue simulating liquids
Conductivity	0,1 ... 17 S/m	$\epsilon'' < 1000$		
		20 MHz	4,8 %	Head and Body tissue
		30 MHz ... 3 GHz	3,7 %	simulating liquids
		3 GHz ... 6 GHz	6,7 %	
		20 MHz ... 1 GHz	7,4 %	Aqueous NaCl solutions with $\epsilon'' < 1000$
		1 GHz ... 6 GHz	2,3 %	Moderate Loss polar liquids, e.g. Ethanol, Methanol, Ethanediol
Loss Tangent	0 ... 1,2	1 GHz ... 6 GHz	8,7 %	Liquids
		20 MHz ... 65 MHz	0,07	Polar liquids
		65 MHz ... 300 MHz	0,02	
		300 MHz ... 1 GHz	0,05	Liquids
		300 MHz ... 6 GHz	0,015	Planar low loss substrate, e.g. permittivity = 10

¹ Slightly depending on the frequency and probe type